

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

OCT 26 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130144 ✓

Name Lewis, Lonnie, B

ID# 7130 ✓ Date 14 Oct 2010 ✓

A

Agency Elmendorf AFB

Phone # 552-4661

Instrument Location Bldg 8517 20th Street Elmendorf AFB, AK

B

Alco S/N X 124658 ✓

Target Value .081 ✓

High Pressure 325

Alco Test Value Average .087 ✓
1st Alco

.088 ✓
2nd Alco

Signature [Signature]

(OVER)

OMB
12/10/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of DEC, 2010.

[Signature]
Daniel L. Nicholas

Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130144

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

OCTOBER 14, 2010

OPERATOR'S NAME:

LEWIS LONNIE B

OPERATOR'S NUMBER: 7130

SUBJECT'S LAST NAME:

NA

SUBJECT'S FIRST NAME/MI :

NA

O.J.L. #: VERIFICATION OF CAL

DEPT/AGENCY: EDF4

CASE/REPORT: NA

TEST TYPE: U

ALCO TARGET VALUE: .001

ALCO S/N: X124658

— BREATH ANALYSIS —

.001 ADJUSTED FOR 29.70 in
ALCO TARGET .000 15:41
BLANK TEST .000 15:42
INTERNAL STANDARD VERIFIED 15:43
ALCO TU 29.70 in .007 15:43
BLANK TEST .000 15:44
SUBJECT SAMPLE .000 15:44
BLANK TEST .000 15:45
ALCO TU 29.70 in .008 15:45
BLANK TEST .000 15:46

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

OCTOBER 14, 2010

TIME 15:29

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/00/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 49c

BAROMETER: 29.70 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~

Pressure 32.5 PSI